



A.B.N. 28 006 859 856
Phone: 61-3-9729 0333
Fax: 61-3-9761 7878
Email: info@ore.com.au
Web: www.ore.com.au

CERTIFICATE OF ANALYSIS FOR
COPPER ORE REFERENCE MATERIAL
OREAS 902

Table 1. 4-Acid ICP-OES/MS - Certified Values, SDs, 95% Confidence and Tolerance Limits for OREAS 902

Constituent	Certified Value	1SD	95% Confidence Limits		95% Tolerance Limits	
			Low	High	Low	High
Four Acid Digestion ICP-OES/MS						
Aluminium, Al (wt.%)	4.74	0.219	4.64	4.84	4.63	4.85
Antimony, Sb (ppm)	1.65	0.086	1.61	1.69	1.58	1.72
Arsenic, As (ppm)	574	24.6	562	586	557	591
Barium, Ba (ppm)	170	6.3	167	173	165	176
Beryllium, Be (ppm)	2.23	0.195	2.13	2.33	2.12	2.34
Bismuth, Bi (ppm)	8.49	0.362	8.33	8.66	8.22	8.77
Calcium, Ca (wt.%)	4.05	0.142	3.98	4.13	3.95	4.16
Cerium, Ce (ppm)	75	3.9	72	77	72	78
Cesium, Cs (ppm)	2.88	0.096	2.82	2.93	2.76	2.99
Chromium, Cr (ppm)	51	6	47	54	47	54
Cobalt, Co (ppm)	926	27.8	912	940	901	951
Copper, Cu (wt.%)	0.301	0.008	0.298	0.304	0.293	0.309
Gallium, Ga (ppm)	11.7	0.63	11.3	12.1	11.4	12.1
Germanium, Ge (ppm)	0.18	0.03	0.15	0.21	IND	IND
Hafnium, Hf (ppm)	4.43	0.316	4.24	4.62	4.33	4.54
Indium, In (ppm)	0.25	0.03	0.23	0.27	0.24	0.26
Iron, Fe (wt.%)	3.19	0.144	3.12	3.27	3.11	3.28
Lanthanum, La (ppm)	36.7	2.41	35.5	37.9	35.2	38.2
Lead, Pb (ppm)	13.3	1.4	12.6	14.0	12.6	14.1
Lithium, Li (ppm)	9.77	0.636	9.44	10.11	9.29	10.26
Lutetium, Lu (ppm)	0.30	0.03	0.28	0.33	0.29	0.32
Magnesium, Mg (wt.%)	2.48	0.096	2.43	2.52	2.42	2.54
Manganese, Mn (wt.%)	0.046	0.003	0.044	0.047	0.044	0.047
Molybdenum, Mo (ppm)	12.2	0.65	11.9	12.5	11.8	12.6
Nickel, Ni (ppm)	164	6.5	161	167	157	171
Phosphorus, P (wt.%)	0.069	0.006	0.066	0.073	0.067	0.072
Potassium, K (wt.%)	3.21	0.168	3.13	3.30	3.10	3.33
Rhenium, Re (ppm)	0.0064	0.0008	0.0058	0.0070	0.0000	0.0000
Rubidium, Rb (ppm)	109	11	103	115	104	113
Scandium, Sc (ppm)	6.9	0.59	6.6	7.2	6.5	7.3
Selenium, Se (ppm)	2.4	0.5	2.1	2.7	IND	IND
Silver, Ag (ppm)	0.343	0.043	0.322	0.363	0.325	0.361
Sodium, Na (wt.%)	0.044	0.007	0.041	0.048	0.043	0.045
Strontium, Sr (ppm)	28.4	1.70	27.6	29.2	27.3	29.5
Sulphur, S (wt.%)	1.76	0.064	1.72	1.79	1.71	1.80
Terbium, Tb (ppm)	0.58	0.07	0.52	0.64	0.56	0.60
Thallium, Tl (ppm)	0.70	0.042	0.68	0.73	0.68	0.73
Thorium, Th (ppm)	11.3	0.67	10.9	11.7	10.9	11.7
Tin, Sn (ppm)	2.05	0.185	1.95	2.15	1.91	2.18
Tungsten, W (ppm)	3.83	0.42	3.60	4.06	3.56	4.11
Uranium, U (ppm)	6.47	0.316	6.28	6.65	6.28	6.65
Vanadium, V (ppm)	54	2.6	53	55	52	56
Ytterbium, Yb (ppm)	1.94	0.128	1.83	2.04	1.77	2.11
Yttrium, Y (ppm)	18.1	1.41	17.4	18.9	17.6	18.6
Zirconium, Zr (ppm)	150	11.5	144	155	145	155

Note: intervals may appear asymmetric due to rounding.

Table 2. Aqua Regia ICP-OES/MS - Certified Values, SDs, 95% Confidence and Tolerance Limits for OREAS 902

Constituent	Certified Value	1SD	95% Confidence Limits		95% Tolerance Limits	
			Low	High	Low	High
Aqua Regia Digestion ICP-OES/MS						
Aluminium, Al (wt.%)	0.535	0.056	0.501	0.569	0.512	0.558
Antimony, Sb (ppm)	0.89	0.10	0.83	0.95	0.85	0.92
Arsenic, As (ppm)	569	32.7	555	584	552	586
Beryllium, Be (ppm)	0.99	0.074	0.95	1.04	0.91	1.08
Bismuth, Bi (ppm)	8.43	0.766	8.05	8.81	8.23	8.63
Calcium, Ca (wt.%)	4.19	0.286	4.05	4.33	4.10	4.28
Cerium, Ce (ppm)	28.3	3.7	25.8	30.8	26.9	29.7
Cesium, Cs (ppm)	0.30	0.03	0.28	0.32	0.27	0.32
Chromium, Cr (ppm)	24.1	2.8	22.8	25.4	22.3	25.9
Cobalt, Co (ppm)	908	67.2	875	941	885	930
Copper, Cu (wt.%)	0.308	0.012	0.303	0.314	0.302	0.315
Gallium, Ga (ppm)	1.55	0.105	1.48	1.63	1.47	1.64
Hafnium, Hf (ppm)	0.54	0.047	0.51	0.57	0.52	0.57
Indium, In (ppm)	0.22	0.018	0.20	0.23	0.21	0.22
Iron, Fe (wt.%)	3.04	0.176	2.94	3.13	2.97	3.10
Lead, Pb (ppm)	10.7	0.98	10.3	11.2	10.3	11.1
Lithium, Li (ppm)	3.7	0.6	3.3	4.1	3.4	4.0
Lutetium, Lu (ppm)	0.088	0.010	0.078	0.099	IND	IND
Magnesium, Mg (wt.%)	2.24	0.120	2.18	2.31	2.20	2.29
Manganese, Mn (wt.%)	0.046	0.002	0.045	0.048	0.045	0.047
Molybdenum, Mo (ppm)	12.6	0.92	12.1	13.1	12.2	13.0
Nickel, Ni (ppm)	159	9.4	154	164	155	163
Phosphorus, P (wt.%)	0.067	0.003	0.066	0.069	0.065	0.069
Potassium, K (wt.%)	0.268	0.032	0.249	0.287	0.253	0.283
Rhenium, Re (ppm)	0.0062	0.0006	0.0059	0.0065	0.0000	0.0000
Rubidium, Rb (ppm)	9.9	1.1	9.2	10.6	9.5	10.3
Scandium, Sc (ppm)	2.93	0.36	2.76	3.10	2.78	3.08
Selenium, Se (ppm)	1.94	0.37	1.74	2.15	1.80	2.08
Silver, Ag (ppm)	0.284	0.031	0.271	0.297	0.265	0.303
Strontium, Sr (ppm)	21.8	1.55	21.0	22.6	21.0	22.6
Sulphur, S (wt.%)	1.78	0.142	1.70	1.85	1.73	1.82
Terbium, Tb (ppm)	0.31	0.03	0.28	0.35	0.29	0.33
Thallium, Tl (ppm)	0.24	0.03	0.22	0.26	0.23	0.25
Thorium, Th (ppm)	5.0	0.47	4.8	5.3	4.9	5.2
Uranium, U (ppm)	2.20	0.26	2.05	2.36	2.14	2.26
Vanadium, V (ppm)	8.79	1.54	7.88	9.70	IND	IND
Ytterbium, Yb (ppm)	0.61	0.054	0.56	0.66	IND	IND
Yttrium, Y (ppm)	7.9	0.68	7.5	8.4	7.7	8.2
Zirconium, Zr (ppm)	15.0	2.5	13.4	16.5	14.2	15.8

Note: intervals may appear asymmetric due to rounding.

Table 3. Copper Soluble - Certified Values, SDs, 95% Confidence and Tolerance Limits for OREAS 902

Constituent	Certified Value	1SD	95% Confidence Limits		95% Tolerance Limits	
			Low	High	Low	High
Acid Leach 5% H₂SO₄						
Copper Soluble, Cu-Sol (wt.%)	0.111	0.011	0.105	0.118	0.109	0.114

Note: intervals may appear asymmetric due to rounding.

Table 4. Indicative Values for OREAS 902

Constituent	Unit	Value	Constituent	Unit	Value	Constituent	Unit	Value
Four Acid Digestion ICP-OES/MS								
Cd	ppm	0.03	Ho	ppm	0.6	Ta	ppm	0.5
Dy	ppm	3.2	Nb	ppm	5.9	Te	ppm	0.03
Er	ppm	2.0	Nd	ppm	29	Ti	wt.%	0.16
Eu	ppm	1.0	Pr	ppm	7.9	Tm	ppm	0.3
Gd	ppm	4.1	Sm	ppm	5.4	Zn	ppm	7.6
Aqua Regia Digestion ICP-OES/MS								
Au	ppm	0.003	Ge	ppm	0.07	Sm	ppm	2.7
B	ppm	13	Hg	ppm	0.05	Sn	ppm	0.4
Ba	ppm	14	Ho	ppm	0.3	Ta	ppm	< 0.01
Cd	ppm	0.02	La	ppm	13	Te	ppm	0.03
Dy	ppm	1.4	Na	wt.%	0.01	Ti	wt.%	0.003
Er	ppm	0.7	Nb	ppm	0.05	Tm	ppm	0.09
Eu	ppm	0.5	Nd	ppm	12	W	ppm	2.1
Gd	ppm	2.3	Pr	ppm	2.7	Zn	ppm	5.6
Fire Assay								
Au	ppb	50						

INTRODUCTION

OREAS reference materials are intended to provide a low cost method of evaluating and improving the quality of analysis of geological samples. To the geologist they provide a means of implementing quality control in analytical data sets generated in exploration from the grass roots level through to prospect evaluation, and in grade control at mining operations. To the analyst they provide an effective means of calibrating analytical equipment, assessing new techniques and routinely monitoring in-house procedures.

SOURCE MATERIALS

OREAS 902 is a low grade transitional copper ore certified reference material. It is one of a suite of four transitional to oxide copper CRMs prepared from CST's Lady Annie Mine, located 120 kms northwest of Mount Isa, Queensland, Australia. Mineralisation at Lady Annie is hosted in dolomitic, carbonaceous and argillaceous sandstones and siltstones. The oxide deposits consist primarily of near surface malachite mineralisation with minor cuprite, chrysocolla and chalcocite extending from surface to a depth of 60 to 100 m. The oxide copper deposit is underlain by deeper transition and sulphide mineralisation. The

primary copper sulphide mineralisation at depth is dominated by chalcocite and chalcopyrite and appears to be structurally controlled, being commonly associated with fault-related silicification.

COMMUNITION AND HOMOGENISATION PROCEDURES

The material constituting OREAS 902 was prepared in the following manner:

- drying to constant mass at 105°C;
- crushing;
- milling to 100% minus 30 microns;
- homogenisation;
- packaging into 10g units in laminated foil pouches and into 1kg units in plastic jars.

ANALYTICAL PROGRAM

Nineteen commercial analytical laboratories participated in the program to characterise the elements reported in Tables 1 to 4. The following methods were employed:

- Four acid digestion with ICP-OES and ICP-MS finish (18 laboratories)
- Aqua regia digestion with ICP-OES and ICP-MS finish (19 laboratories)
- 5% H₂SO₄ acid leach with AAS or ICP-OES finish (14 laboratories)

For the round robin program twenty 1kg test units were taken at predetermined intervals during the bagging stage after final blending and are considered representative of the entire batch. The six samples received by each laboratory were obtained by taking two 110g scoop splits from each of three separate 1kg test units. This format enabled nested ANOVA treatment of the results to evaluate homogeneity.

Results, together with uncorrected means, medians, standard deviations, relative standard deviations and percent deviation of lab means from the corrected mean of means (PDM³) are presented in the detailed certification report for this CRM (Hamlyn, 2011).

STATISTICAL ANALYSIS

Certified Values, Standard Deviations, Confidence and Tolerance Limits have been determined for each analytical method following removal of individual and laboratory outliers (see Tables 1-3). Certified Values are the mean of means after outlier filtering. The 95% Confidence Limit is a measure of the reliability of the certified value, i.e. the narrower the Confidence Interval the greater the certainty in the Certified Value. It should not be used as a control limit for laboratory performance.

Indicative values (Table 4) are provided where i) the number of laboratories reporting a particular analyte is insufficient (< 5) to support certification; ii) interlaboratory consensus is poor; or iii) a significant proportion of results are outlying or reported as less than detection limits.

Standard Deviation values (1SDs) are reported in Tables 1-3 and provide an indication of a level of performance that might reasonably be expected from a laboratory being monitored by this CRM in a QA/QC program. They take into account errors attributable to measurement uncertainty and CRM variability. For an effective CRM the contribution of the latter should be negligible in comparison to measurement errors. The Standard Deviation values include all sources of measurement uncertainty: between-lab variance, within-run variance (precision errors) and CRM variability. The SD for each analyte's certified value is calculated from the same filtered data set used to determine the certified value, i.e. after removal of all individual, lab dataset (batch) and 3SD outliers (single iteration). These outliers can only be removed after the absolute homogeneity of the CRM has been independently established, i.e. the outliers must be confidently deemed to be analytical rather than arising from inhomogeneity of the CRM. The standard deviation is then calculated for each analyte from the pooled accepted analyses generated from the certification program.

As a guide two or more analytical results lying outside the 2SD window may be regarded as warning or rejection, and rejection for single results lying outside the 3SD window in QC monitoring, although their precise application should be at the discretion of the QC manager concerned.

Tolerance Limits (ISO Guide 3207) were determined using an analysis of precision errors method and are considered a conservative estimate of true homogeneity. The meaning of tolerance limits may be illustrated for copper by 4-acid digestion, where 99% of the time ($1-\alpha=0.99$) at least 95% of subsamples ($\rho=0.95$) will have concentrations lying between 0.293 and 0.309 wt.%. Put more precisely, this means that if the same number of subsamples were taken and analysed in the same manner repeatedly, 99% of the tolerance intervals so constructed would cover at least 95% of the total population, and 1% of the tolerance intervals would cover less than 95% of the total population (ISO Guide 35).

The homogeneity of OREAS 902 has also been evaluated in an ANOVA study for all certified analytes. This study indicates no evidence that between-unit variance is greater than within-unit variance.

Based on the statistical analysis of the results of the interlaboratory certification program it can be concluded that OREAS 902 is fit-for-purpose as a certified reference material (see 'Intended Use' below).

A detailed report covering statistical treatment and tabulation of the analytical results is available on request as a separate pdf document (Certification Report for OREAS 902).

PREPARER AND SUPPLIER OF THE REFERENCE MATERIAL

Low grade transitional-sulphide copper ore reference material OREAS 902 has been prepared, certified and is supplied by:

ORE Research & Exploration Pty Ltd
6-8 Gatwick Road
Bayswater North VIC 3153
AUSTRALIA

Tel: +613-9729 0333
Fax: +613-9761 7878
Web: www.ore.com.au
Email: info@ore.com.au

OREAS 902 has been packaged in single-use laminated foil pouches in 10g units. 1kg units in plastic jars are also available upon request.

INTENDED USE

OREAS 902 is intended for the following uses:

- for the monitoring of laboratory performance in the analysis of analytes reported in Tables 1-3 in geological samples
- for the verification of analytical methods for analytes reported in Tables 1-3
- for the calibration of instruments used in the determination of the concentration of analytes reported in Tables 1-3

STABILITY AND STORAGE INSTRUCTIONS

OREAS 902 has been sourced from low grade transitional-sulphide copper ore. In its unopened state and under normal conditions of storage it has a shelf life beyond five years. Its stability will be monitored at regular intervals and purchasers notified if any changes are observed.

INSTRUCTIONS FOR THE CORRECT USE OF THE REFERENCE MATERIAL

The certified values refer to the concentration level of analytes in their packaged state. The CRM should therefore not be dried prior to weighing and analysis.

HANDLING INSTRUCTIONS

Fine powders pose a risk to eyes and lungs and therefore standard precautions such as the use of safety glasses and dust masks are advised.

LEGAL NOTICE

Ore Research & Exploration Pty Ltd has prepared and statistically evaluated the property values of this reference material to the best of its ability. The Purchaser by receipt hereof releases and indemnifies Ore Research & Exploration Pty Ltd from and against all liability and costs arising from the use of this material and information.

CERTIFYING OFFICER

Craig Hamlyn (B.Sc. Hons - Geology), Technical Manager - ORE

PARTICIPATING LABORATORIES

Acme Analytical Laboratories, Vancouver, BC, Canada
Activation Laboratories, Ancaster, Ontario, Canada
Activation Laboratories, Thunder Bay, Ontario, Canada
ALS, Brisbane, QLD, Australia
ALS, Callao, Lima, Peru

ALS, Johannesburg, Gauteng, South Africa
ALS, La Serena, Coquimbo, Chile
ALS, Perth, WA, Australia
ALS, Vancouver, BC, Canada
BV Amdel, Adelaide, SA, Australia
BV Ultra Trace, Perth, WA, Australia
Intertek Genalysis, Perth, WA, Australia
McPhar Geoservices (Phil) Inc., Manila, Philippines
SGS Mineral Services, Lakefield, Ontario, Canada
SGS Mineral Services, Perth, WA, Australia
SGS Mineral Services, Toronto, Ontario, Canada
SGS Mineral Services, Townsville, QLD, Australia
SGS Mineral Services, Vancouver, BC, Canada
Zarazma Mineral Studies, Tehran, Iran

REFERENCES

- ISO Guide 35 (2006), Certification of reference materials - General and statistical principals.
ISO Guide 3207 (1975), Statistical interpretation of data - Determination of a statistical tolerance interval.
Hamlyn, C. L. (2011), Certification Report for OREAS 902.